



	<b>Fibrinogen test</b>					
<b>Application</b>	For the quantitative estimation of Fibrinogen in plasma					
<b>Reagent</b>	<ul style="list-style-type: none"> <li>• Lyophilised Thrombin ~ 50 NIH units</li> <li>• Lyophilised Human Fibrinogen calibrator</li> <li>• Ready to use Owrens Buffer</li> </ul>					
<b>Source</b>	Bovine / Human Plasma					
<b>Normal Fibrinogen Values</b>	150 - 400 mg/dl					
<b>Storage / Stability</b>	2-8°C, 18 months					
<b>In Use Stability</b>	Reconstituted, <table border="1" data-bbox="663 1384 1343 1559"> <tr> <td data-bbox="663 1384 1007 1473">           Thrombin reagent         </td> <td data-bbox="1007 1384 1343 1473">           2-8° C, 30 days 20-25° C, 4 hours         </td> </tr> <tr> <td data-bbox="663 1473 1007 1559">           Calibrator         </td> <td data-bbox="1007 1473 1343 1559">           2-8° C, 6 hours 20-25° C, 2 hours         </td> </tr> </table>		Thrombin reagent	2-8° C, 30 days 20-25° C, 4 hours	Calibrator	2-8° C, 6 hours 20-25° C, 2 hours
Thrombin reagent	2-8° C, 30 days 20-25° C, 4 hours					
Calibrator	2-8° C, 6 hours 20-25° C, 2 hours					

Presentation	Pack	Cat.No.
FIBROQUANT®	20 Tests	10641020

