



## Dilute Russell's Viper Venom Time (dRVVT) based Lupus Anticoagulant detection system

<b>Application</b>	For screening and confirmation of Lupus Anticoagulants in plasma	
<b>Sensitivity</b>	Phospholipid dependant Lupus Anticoagulants	
<b>Specificity</b>	Not affected by contact and antihaemophilic factor deficiencies or corresponding inhibitors	
<b>Reagents</b>	• Screening Reagent	Phospholipids Lyophilised with dRVV
	• Confirmation Reagent	Enriched Phospholipids Lyophilised with dRVV
	• Calcium Chloride	Ready to use 0.025M solution, insensitive to heparin upto 0.4 IU/ml
<b>Normal Range</b>	<ul style="list-style-type: none"> <li>• With Screening Reagent 28 - 45 seconds</li> <li>• With Confirmation Reagent, 28 - 40 seconds</li> <li>• Results may also be expressed as Ratio</li> </ul>	
<b>Storage / Stability</b>	2-8°C, 12 months	
<b>In Use Stability</b>	Reconstituted stability 2-8°C, 5 days for screening and confirmation reagent	

Presentation	Pack	Cat.No.
LA Confirm	5 Tests	10673005