IMPORTANT NOTE

- 1. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbance readings.
- It is recommended to use the multiple channel pipettes to avoid time effect. A full plate of 96 wells may be used if automated pipetting is available.
- 3. Duplication of standards & samples is not mandatory but may provide information on reproducibility & application errors.

LIMITATIONS OF THE ASSAY

- 1. As with all diagnostic tests, a definite clinical diagnosis should not be based on the results of a single test, but should only be made by the physician after all clinical and laboratory findings have been evaluated.
- The activity of the enzyme used is temperature-dependent and the OD values may vary. The higher the room temperature
 (+18°C to +25°C) during substrate incubation, the greater will be the OD values. Corresponding variations apply also to
 the incubation times. However, the standards are subject to the same influences, with the result that such variations will
 be largely compensated in the calculation of the result.
- Adaptation of this assay for use with automated sample processors and other liquid handling devices, in whole or in part, may yield differences in test results from those obtained using the manual procedure. It is the responsibility of each laboratory to validate that their automated procedure yields test results within acceptable limits.
- Insufficient washing (e.g., less than 5 wash cycles, too small wash buffer volumes, or shortened reaction times) can lead
 to incorrect OD values.

BIBLIOGRAPHY

- Gold P, Freedman S O. Demonstration of tumor specific antigen in human colonic carcinomata by immunologic tolerance and absorption techniques. J Exp Med 1965;127:439-462.
- Thompson D P M, Krupey J, Freedman S O, et al. The radioimmunoassay of circulating carcinoembryonic antigen of the human digestive system. Proc Natl Acad Sci USA 1969:64:161-167.
- Schwartz M K. Tumor markers in diagnosis and screening. In: Ting S W, Chen J S, Schwartz M K, eds. Human tumor markers. Amsterdam: Elsevier Science. 1987:3-16.
- Zamcheck N. and Martin E.W. Sequential Carcinoembryonic Antigen Levels in Pancreatic Cancer: Some Clinical Correlations. Cancer 1981:47:1620-1627.
- Mughal A.W., Hortobagyi G. N., Fritsche H.A., Buzdar A.U. Yap H-Y., and Blumenschein G.R. Serial Plasma Carcinoembryonic Antigen Measurements During Treatment of Metastatic Breast Cancer. JAMA 1983; 259:1881-1886.

SYMBOL KEYS

Temperature Limitation	Consult Instructions for use	Date of Manufacture	LOT Batch Number / Lot Number
Manufacturer	In vitro Diagnostic Medical Device	This side up	Σ Contains sufficient
Use by	REF Catalogue Number	Do not reuse	for <n> tests</n>



Manufactured by

Zephyr Biomedicals

A Division of Tulip Diagnostics (P) Ltd.

M 46-47, Phase III B, Verna Industrial Estate, Verna, Goa - 403 722, INDIA.

Regd. Office: Gitanjali, Tulip Block, Dr. Antonio Do Rego Bagh, Alto Santacruz,
Bambolim Complex P.O., Goa - 403 202, INDIA.



Enzyme Linked Immunosorbent assay for the Quantitative Determination of Carcinoembryonic Antigen (CEA) in Human Serum

FOR IN VITRO DIAGNOSTIC USE ONLY

Store at 2°C to 8°C

INTENDED LISE

Qualisa™ CEA Sandwich ELISA test is intended for the quantitative determination of Carcinoembryonic Antigen (CEA) in human serum. For In Vitro Diagnostic Use only.

INTRODUCTION

Carcinoembryonic antigen (CEA) is a cell-surface 200-kd glycoprotein. In 1969, it was reported that plasma CEA was elevated in 35 of 36 patients with adenocarcinoma of the colon and that CEA titers decreased after successful surgery. Normal levels were observed in all patients with other forms of cancer or benign diseases. Subsequent studies have not confirmed these initial findings, and it is now understood that elevated levels of CEA are found in many cancers. Increased levels of CEA are observed in more than 30% of patients with cancer of the lung, liver, pancreas, breast, colon, head or neck, bladder, cervix and prostate. Elevated plasma levels are related to the stage and extent of the disease, the degree of differentiation of the tumor, and the site of metastasis. CEA is also found in normal tissue.

PRINCIPLE

Qualisa™ CEA Quantitative Test Kit is a sandwich-based enzyme-linked immunosorbent assay. The assay system utilizes one monoclonal anti-CEA antibody for solid phase (microtiter wells) immobilization and another mouse monoclonal anti-CEA antibody in the antibody-enzyme (horseradish peroxidase) conjugate solution. The test sample is allowed to react simultaneously with the two antibodies, resulting in the CEA molecules being sandwiched between the solid phase and enzyme-linked antibodies. After incubation the wells are washed and bound enzyme is detected by adding substrate. The reaction is stopped after specified time with stop solution and absorbance is determined for each well using an ELISA reader. The concentration of CEA is directly proportional to the color intensity of the test sample.

MATERIALS AND COMPONENTS

Materials provided with the test kits:

- Coated Microwells: Microwells coated with monoclonal anti- CEA antibody.
- CEA Enzyme Conjugate. Ready to use.
- TMB Substrate. Ready to use.
- Stop Solution. Ready to use.
- CEA Standard set of 6 standards labeled as A to F in liquid form. For standard Concentrations refer vial label.
- Wash Buffer Concentrate (20X).

Materials required but not provided

- Precision pipettes: 10-100µl, 20-200µl, 100-1000µl
- Disposable pipette tips
- Distilled water
- Disposable Gloves
- ELISA reader
- ELISA washer

STORAGE AND STABILITY

- Qualisa[™] CEA kit is stable at 2-8°C upto expiry date printed on the label.
- Coated microwells should be used within one month upon opening the pouch provided that once opened, the pouch must be resealed to protect from moisture. If the colour of the dessicant has changed from blue to pink at the time of opening the pouch, another coated microwells pouch should be used.
- 3. Diluted Wash Buffer is stable upto one week when stored at 2-8°C.

SPECIMEN COLLECTION

- 1. Collect Blood specimen by venipuncture according to the standard procedure.
- 2. Only serum should be used.
- 3. Avoid grossly hemolytic, lipemic or turbid samples.
- 4. Preferably use fresh samples. However, specimens can be stored up to 48 hours at 2-8°C, for short duration.
- 5. For longer storage, specimens can be frozen at -20°C. Thawed samples must be mixed prior to testing.
- Do not heat inactivate before use.
- 7. Specimen containing precipitate or particulate matter should be clarified by centrifugation prior to use.
- 8. Specimen should be free from particulate matter and microbial contamination.

PRECAUTIONS


- 1. Bring all reagents and specimen to room temperature before use.
- 2. Do not pipette any material by mouth.
- 3. Do not eat, drink or smoke in the area where testing is done.
- 4. Use protective clothing and wear gloves when handling samples.
- 5. Use absorbent sheet to cover the working area.
- 6. Immediately clean up any spills with sodium hypochlorite.
- 7. All specimens and standards should be considered potentially infectious and discarded appropriately.
- 8. Neutralize acid containing waste before adding hypochlorite.
- 9. Do not use kit after the expiry date.
- 10. Do not mix components of one kit with another.
- 11. Always use new tip for each specimen and reagent.
- 12. Do not allow liquid from one well to mix with other wells.
- 13. Do not let the strips dry in between the steps.

REAGENT PREPARATION

- All reagents should be brought to room temperature (18-25°C) and mixed by gently inverting or swirling prior to use. Do not induce foaming.
- 2. Dilute wash buffer 20 times (for example add 5ml concentrated buffer to 95 ml distilled or deionized water) Mix well before use.

TEST PROCEDURE

- 2. Dispense **50 μI** of enzyme conjugate reagent into each well. Incubate at room temperature (18-25°C), for **45 minutes**.
- 3. After incubation, empty the microtitre wells and wash the plate 5 times with 350µl of diluted wash buffer. Strike the microtitre plate sharply onto absorbent paper towel to remove all residual droplets.
- 4. Dispense 100 µl of TMB substrate into each well. Incubate at room temperature(18-25°C) in the dark, for 20 minutes.
- Stop the reaction by adding 100 µl of Stop Solution to each well. Gently mix for 10 seconds until the blue color completely changes to vellow.
- 6. Read the optical density at 450/630 nm with a microtiter plate reader within 15 minutes.



CALCULATION OF RESULTS


Construct a standard curve by plotting the absorbance obtained from each reference standards against its concentration in ng/ml on the graph paper, with absorbance values on the vertical or Y axis and concentrations on the horizontal or X axis. Use the absorbance values for each specimen to determine the corresponding concentration of CEA in ng/ml from the standard curve. Any diluted specimens must be corrected by the appropriate dilution factor.

Example of Standard curve

Results of a typical standard run with optical density reading at 450nm (ref 600-700nm) shown in the Y axis against CEA concentrations shown in the X axis.

Suggest: Use 4-Parameter Standard curve to calculate sample values.

CEA (ng/ml)	Absorbance
Α	0.0094
В	0.1315
С	0.4984
D	1.0172
E	1.6667
F	1.9562



This standard curve is for the purpose of illustration only and should not be used to calculate unknowns. Each user should obtain his or her own standard curve and data.

Expected values and sensitivity

The most complete study of CEA is a compilation of collaborative studies in which CEA values in 35,000 samples from more than 10,000 patients and controls were analyzed. Of 1425 normal persons who did not smoke, 98.7% had values less than 5.0ng/ml. It is recommended that each laboratory establish its own normal range. The minimum detectable concentration of CEA by this assay is estimated to be 1.0 ng/ml.

PERFORMANCE CHARACTERISTICS

A) Internal Evaluation:

Accuracy: In an internal study Qualisa[™] CEA was evaluated against commercially available licensed kit with 89 random clinical samples, & Qualisa[™] CEA has demonstrated 100% clinical correlation with the commercially available licensed kit

Precision: Qualisa[™] CEA was evaluated with licensed external Quality controls for Precision Studies & following is the data:

Controls	No. of testings	Mean Control values with Qualisa™ CEA	Coefficient of Variation (CV)
Level 1	10	2.02	2.94
Level 2	10	24.02	1.32
Level 3	10	106.42	1.59

B) External Evaluation:

Qualisa™ CEA ELISA has been evaluated by a NABL accredited lab against their reference method. In this evaluation Qualisa™ CEA ELISA has demonstrated 100% correlation with the reference method.

*Data file: Zephyr Biomedicals (A Division of Tulip Diagnostics Pvt. Ltd.).