



Zephyr Biomedicals

Qualisa™ AMH Enzyme linked Immunosorbent Assay (ELISA) for Quantitative Determination of Anti-Mullerian Hormone (AMH) in Human Serum

Intended Use: _____

Qualisa™ AMH ELISA test is intended for the quantitative determination of Anti-Mullerian Hormone (AMH) in Human Serum.

Qualisa™ AMH kit Components:

Coated Microwells	Coated Microwells in 3x8 microwells strips/pouch.
Conjugate	HRP Conjugate concentrate (20X).
Conjugate Diluent	Diluent for conjugate dilution.
Standards	Standards with known concentration for calibration.
Wash Buffer	Buffer contains surfactants (20X). To be diluted 20 times with distilled or deionized water.
Substrate	Solution containing Tetramethyl benzidine (TMB) and hydrogen peroxide. Ready to use.
Stop Solution	Diluted Sulphuric acid. Ready to use.
Other accessories	Microwell holder, ELISA protocol sheet, Plate sealer and Package Insert.

Performance Characteristics: Evaluated by a NABL accredited lab

Excellent Precision: CV < 10%

FEATURES	BENEFITS
Shorter Turn Around Time (TAT)	Faster reporting time
Stable Calibration	Calibration stability up to 45 days
Unique packaging of microwells (3x8 microwell strip/pouch)	Reduces multiple exposure of microwells

Storage / Stability	Temperature	Duration
Unopened kit	2-8°C	12 Months

Available Pack Sizes
96 Tests

