

	<p><b>Chemiluminescence Assay for the Quantitative Determination of Carcinoembryonic Antigen (CEA) in Human Serum.</b></p> <p><b>FOR IN VITRO DIAGNOSTIC USE ONLY</b></p>
<p><b>Application</b></p>	<p>For the quantitative determination of Carcinoembryonic Antigen (CEA) in Human Serum.</p>
<p><b>Principle</b></p>	<p>Chemiluminescence (CLIA)</p>
<p><b>Reagents</b></p>	<ul style="list-style-type: none"> <li>• <b>Coated microwells:</b>96 microwells (3x8 wells x 4) are coated with Anti- CEA antibody.</li> <li>• <b>Standards:</b> CEA standards with desired concentration for calibration.</li> <li>• <b>CEA Enzyme Conjugate.</b> Ready to use.</li> <li>• <b>Substrate A:</b> Chemiluminescent substrate containing enhanced luminol solution.</li> <li>• <b>Substrate B:</b> Chemiluminescent substrate containing stabilized peroxide solution.</li> <li>• <b>Wash buffer:</b> Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.</li> </ul>
<p><b>Test duration</b></p>	<p>40 minutes</p>
<p><b>Storage / Stability</b></p>	<p>2-8°C, 12 months</p>

Presentation	Pack	Cat.No.
<p><b>Electra CEA</b> kit contains:</p> <ul style="list-style-type: none"> <li>• Microwell strips (3 x 8 wells/pouch)</li> <li>• Reagents</li> <li>• Accessories</li> <li>• Standards</li> </ul>	<p>96 Tests</p>	<p>513050096</p>

