



	Chemiluminescence Assay for Qualitative Detection of Cytomegalovirus IgM antibody in Human Serum
Application	For the Qualitative detection of IgM antibodies to cytomegalovirus (CMV) infection in human serum
Principle	Chemiluminescence (CLIA)
Reagents	<ul style="list-style-type: none"> • Coated microwells: 96 microwells (3x8 wells x 4) Purified anti-human IgM agglutinating sera coated wells. • Sample Diluent. Ready to use. • HRPO Enzyme Conjugate: Ready to use. • Negative Control: Ready to use. • Positive Control: Ready to use. • Calibrator: Ready to use • Substrate A: Chemiluminescent substrate containing enhanced luminol solution. • Substrate B: Chemiluminescent substrate containing stabilized peroxide solution. • Wash buffer: Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.
Test duration	100 minutes
Storage / Stability	2-8°C, 12 months

Presentation	Pack	Cat.No.
Electra CMV IgM kit contains: <ul style="list-style-type: none"> • Microwell strips (3 x 8 wells/pouch) • Reagents • Accessories 	96 Tests	512100096

