



 	Chemiluminescence Assay for Qualitative Detection of HSV 1, 2 IgM antibody in Human Serum
Application	For the Qualitative detection of IgM antibodies herpes simplex virus (HSV) infection, or for evaluating paired sera for the presence of a significant increase in herpes specific IgM.
Principle	Chemiluminescence (CLIA)
Reagents	<ul style="list-style-type: none"> • Coated microwells: 96 microwells (3x8 wells x 4) Purified HSV 1,2 antigen coated wells. • Sample Diluent: Ready to use. • HRPO Enzyme Conjugate: Ready to use. • Negative Control: Ready to use. • Positive Control: Ready to use. • Calibrator: Ready to use • Substrate A: Chemiluminescent substrate containing enhanced luminol solution. • Substrate B: Chemiluminescent substrate containing stabilized peroxide solution. • Wash buffer: Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.
Test duration	100 minutes
Storage / Stability	2-8°C, 12 months

Presentation	Pack	Cat.No.
Electra HSV 1,2 IgM kit contains: <ul style="list-style-type: none"> • Microwell strips (3 x 8 wells/pouch) • Reagents • Accessories 	96 Tests	512130096