



	Chemiluminensence Assay for Quantitative Determination of Progesterone in Human Serum
Application	For the quantitative determination of Progesterone in human serum.
Principle	Chemiluminescence (CLIA)
Reagents	<ul style="list-style-type: none"> • Coated microwells: 96 microwells (3x8 wells x 4) are coated with Goat Anti-Rabbit IgG. • Standards: Progesterone standards with desired concentration for calibrators.. • Rabbit Anti-Progesterone Reagent: Ready to use • Progesterone-HRP Conjugate Concentrate (11X). • Progesterone-HRP Conjugate Diluent. • Substrate A: Chemiluminescent substrate containing enhanced luminol solution. • Substrate B: Chemiluminescent substrate containing stabilized peroxide solution. • Wash buffer: Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.
Test duration	70 minutes
Storage / Stability	2-8°C, 12 months

Presentation	Pack	Cat.No.
Electra PROG kit contains: <ul style="list-style-type: none"> • Microwell strips (3 x 8 wells/pouch) • Reagents • Accessories • Standards 	96 Tests	512050096