



	Chemiluminescence Assay for Qualitative Detection of Rubella IgM antibody in Human Serum.
Application	For the Quantitative detection of IgM antibodies to Rubella virus infection in human serum.
Principle	Chemiluminescence (CLIA)
Reagents	<ul style="list-style-type: none"> • Coated microwells: 96 microwells (3x8 wells x 4) Purified anti-human IgM agglutinating sera coated wells. • Sample Diluent. Ready to use. • HRPO Enzyme Conjugate. Ready to use. • Calibrator. Ready to use • Negative Control. Ready to use. • Positive Control. Ready to use. • Substrate A: Chemiluminescent substrate containing enhanced luminol solution. • Substrate B: Chemiluminescent substrate containing stabilized peroxide solution. • Wash buffer: Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.
Test duration	100 minutes
Storage / Stability	2-8°C, 12 months

Presentation	Pack	Cat.No.
Electra Rubella IgM kit contains: <ul style="list-style-type: none"> • Microwell strips (3 x 8 wells/pouch) • Reagents • Accessories 	96 Tests	512090096

