



	<b>Chemiluminescence Assay for Qualitative Detection of Toxoplasma IgM antibody in Human Serum</b>
<b>Application</b>	For the Qualitative detection of IgM antibodies to Toxoplasma gondii infection in human serum.
<b>Principle</b>	Chemiluminescence (CLIA)
<b>Reagents</b>	<ul style="list-style-type: none"> <li>• <b>Coated microwells:</b> 96 microwells (3x8 wells x 4) Purified anti-human IgM agglutinating sera coated wells.</li> <li>• <b>Sample Diluent.</b> Ready to use.</li> <li>• <b>HRPO Enzyme Conjugate.</b> Ready to use.</li> <li>• <b>Calibrator.</b> Ready to use</li> <li>• <b>Negative Control.</b> Ready to use.</li> <li>• <b>Positive Control.</b> Ready to use.</li> <li>• <b>Substrate A:</b> Chemiluminescent substrate containing enhanced luminol solution.</li> <li>• <b>Substrate B:</b> Chemiluminescent substrate containing stabilized peroxide solution.</li> <li>• <b>Wash buffer:</b> Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.</li> </ul>
<b>Test duration</b>	100 minutes
<b>Storage / Stability</b>	2-8°C, 12 months

Presentation	Pack	Cat.No.
<b>Electra Toxo IgM kit contains:</b> <ul style="list-style-type: none"> <li>• Microwell strips (3 x 8 wells/pouch)</li> <li>• Reagents</li> <li>• Accessories</li> </ul>	96 Tests	512080096

