

	Enzyme Linked Immunosorbent assay for the Quantitative Determination of Carcinoembryonic Antigen (CEA) in Human Serum
Application	For the quantitative determination of Carcinoembryonic Antigen (CEA) in Human Serum
Principle	Enzyme Linked Immunosorbent Assay (ELISA)
Reagents	<ul style="list-style-type: none"> • Coated microwells: 48 microwells (3x8 wells x 2) are coated with Anti-CEA antibody. • Standards: CEA standards with desired concentration for calibration. • CEA Enzyme Conjugate. Ready to use. • Substrate: Solution containing Tetramethyl benzidine (TMB) and hydrogen peroxide. Ready to use. • Wash buffer: Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water. • Stop solution: Diluted Sulphuric acid. Ready to use
Test duration	65 minutes
Storage / Stability	2-8°C, 12 months

Presentation	Pack	Cat.No.
Qualisa CEA kit contains: <ul style="list-style-type: none"> • Microwell strips (3 x 8 wells/pouch) • Reagents • Accessories • Standards 	48 Tests	514050048

