

	Fluorescence Immunoassay for quantitative determination of D-Dimer in human plasma
Application	FIAcheck D-Dimer Fluorescence Immunoassay is intended for the in-vitro quantitative measurement of D-Dimer in human plasma.
Principle	Rapid, lateral flow, quantitative immune-fluorescence assay
Measuring range	0.1-10µg/mL
Reagents	Membrane assembly coated with <ul style="list-style-type: none"> •Immobilized D-Dimer mouse monoclonal antibody1 on test line •Goat anti chicken IgY in control line •Mixture of D-Dimer mouse monoclonal antibody2 labeled with fluorescent microspheres on the binding pad Sample diluent
Test duration	10 mins
Storage / Stability	4-30°C, 12 months

Presentation	Pack	Cat.No.
FIAcheck Anti-CCP kit contains an individually pouched device, sample diluent, calibration QR code card and dilution vials	10 Tests	540010010

