



	<p>Controls for quality control procedures</p>
<p>Application</p>	<p>Turbodyne™ IgE controls are also useful for validating the calibration and performance of the Turbodyne™ IgE reagent using Turbodyne™ SC analyzer.</p>
<p>Reagent</p>	<p>Liquid stable preparation from human serum and fortified with preservatives and stabilizers that are usually not known to interfere in the test results</p>
<p>Storage / Stability</p>	<p>2-8°C, 12 months</p>
<p>In Use Stability</p>	<p>Turbodyne™ IgE ready to use control is stable for 1 week at 2-8°C</p>

Presentation	Pack	Cat.No.
Turbodyne™ IgE control	0.2 ml	108610002

